

Technical Specifications

Republic of the Philippines Department of Health Regional Office 02 Region II Trauma and Medical Center			
Item No. 28	LARYNGOSCOPE, PEDIA	Quantity ABC	2 Units P25,000.00 / Unit
Name of Manufacturer:		Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<ul style="list-style-type: none"> • Handle <ul style="list-style-type: none"> – Solid Metal – Chargeable batteries size AA or size C at the base of handle – Functions with non-rechargeable batteries size AA or size C – Simple recharging in charging base at 220 V – Solid metal contact for activation and illumination – Provide charger for recharging batteries • Blade <ul style="list-style-type: none"> – MAC 1,2,3,4 – Miller 1,2 – High power LED illumination – Reusable • Accessories <ul style="list-style-type: none"> – Charging base at 220V – Storage/Carrying case • Warranty of 2 years • Unit must be available for inspection/demonstration 			

Submitted by:

Name and Signature of the Representative of the Bidder: _____

Name of Company: _____

Date: _____