

# Technical Specifications

Republic of the Philippines  
 Department of Health  
 Regional Office 02  
 Region II Trauma and Medical Center

Item No. 1	<b>Laryngoscope, Pedia</b>	Quantity ABC	2units P25,000.00/unit
Name of Manufacturer:		Country of Origin:	
Brand:		Model:	
<b>PURCHASER'S SPECIFICATION</b>		<b>SUPPLIER'S SPECIFICATION</b>	
<p><i>Specifications:</i></p> <ul style="list-style-type: none"> <li>❖ <b>Handle</b> <ul style="list-style-type: none"> <li>• Solid Metal</li> <li>• Chargeable batteries size AA or size C at the base of handle</li> <li>• Functions with non-rechargeable batteries size AA or size C</li> <li>• Simple recharging in charging base at 220V</li> <li>• Solid metal contact for activation and illumination</li> <li>• Provide charger for recharging batteries</li> </ul> </li> <li>❖ <b>Blade</b> <ul style="list-style-type: none"> <li>• MAC 1,2,3,4</li> <li>• Miller 1,2</li> <li>• High power LED illumination</li> <li>• Reusable</li> </ul> </li> <li>❖ <b>Accessories</b> <ul style="list-style-type: none"> <li>• Charging base at 220V</li> <li>• Storage /Carrying case</li> </ul> </li> <li>❖ <b>Warranty of Two (2) years</b></li> <li>❖ <b>Unit must be available for inspection/demonstration</b></li> </ul>			