

TECHNICAL SPECIFICATIONS

	<i>Lot/Item</i>	<i>Max. Qty</i>		<i>Technical Specifications / Scope of Work</i>	<i>Statement of Compliance</i>
1	Basin, Kidney	50	pcs	plastic, holds atleast 500cc	
2	Blood Transfusion set	800	pcs	w/ G19 needle, transparent flexible and cylindrical drip chamber with 200 micron filter, Regulator clamp to control transfusion rate, Luer lock or rotating Luer lock connector	
3	Catheter, Hemodialysis, curved	4,000	set	Set contains Fr 11-12 x 15-17 cm triple lumen catheter with radiopaque markings and made of medical-grade thermoplastic polyurethane material, with flexible J-tip guidewire in introducer needle, valve cannula G18 x 2 ½ inches, tissue dilator, antimicrobial heparin cap and clamp, trocar, pressure transduction probe, scalpel, syringes, injection needle, suture needle, surgical gloves, sterile dressing	
4	Catheter, hemodialysis, straight	300	pcs	Set contains Fr 12 triple lumen catheter, at least 17 cm in length, with radiopaque markings and made of medical-grade thermoplastic polyurethane material, with flexible J-tip guidewire in advancer, valve cannula G18 x 2 ½ inches, dilator size Fr 11-12, antimicrobial heparin cap and clamp, trocar, pressure transduction probe, scalpel, syringes, injection needle, suture needle, surgical gloves, sterile dressing	
5	Disposable Core Tissue Biopsy Needle	100	pcs	14G x 10cm, length of sample notch 1.9cm, compatible with existing biopsy gun (Bard)	
6	Laminaria 3mm-5mm	600	pcs	3-5mm, sterile, 25pcs/box	
7	Polymer Clip Cartridge, Large	600	pcs	5-13mm, supplier must provide 1 clipper every 50pcs of clips	
8	Polymer Clip Cartridge, Medium	200	pcs	3-10mm, supplier must provide 1 clipper every 50pcs of clips	
9	Suture, Black silk 0 with cutting needle,	16000	pcs	non-absorbable multi-filament braided suture Coated with Silicone or Wax (beeswax/paraffin) At least 75cm length, needle 1/2 circle 25- 31mm, individually packed, sterile packaging	
10	Suture, Balck Silk 6/0	1000	pcs	Black non-absorbable multi-filament braided suture, 45-55 cm length, double arm cutting needle 3/8 circle, 13 mm	
11	Suture, Plain 2/0 round needle	150	pcs	Round needle 3/8 circle, 27-30mm, 70-76 cm length	
12	Suture, Polyglactin 6/0 w/ round needle	500	pcs		
13	Thermal paper for ECG Machine size 210mm x 30M, individually packed	20	pcs		
	ETT introducer (Bougie)	2	box	Latex-free, flexible, blue or yellow color with graduated markings, distal tip bent at a 30 degree angle, vented tip with distal rounded smooth tip, with a TPE conical connector	

14	size 15 Fr, ED 5mm, 70cm length	2	box		
15	size 10Fr, ED 3.3mm, 70cm length	2	box		
16	size 8Fr, ED 2.7mm, 53.5cm length	200	pc		
	ETT introducer (Bougie), medical grade PE	504	bot	Latex-free, flexible, blue or yellow color with graduated markings, distal tip bent at a 30 degree angle, curved tip with distal rounded smooth tip	
17	size 10Fr, ED 3.3mm, 70cm length	30	roll		
18	size 6Fr, ED 2.0mm, 53.5cm length	50	doz		
19	Head Positioner, Neonate	2	doz	Donut type, Gel Material	
20	Head Positioner, Pedia	75	doz	Donut type, Gel Material	
21	Head Positioner, Adult	100	doz	Donut type, Gel Material	
22	Needle, Spinal, Quincke, G26	5	doz	cutting bevel sharp tip, with clear luer lock hub connection, sterile package,	

[Bidders must state in the *Statement of Compliance column* either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution.]

Note:

Presentation/submission of sample/s (properly labelled with the name of company, item # and description) to the BAC Office is required for evaluation purposes. Bidders are required to secure a certification that sample/s was/were presented/submitted for evaluation to be issued by the Technical Working Group.

No.

of samples - 1 piece

Name of Company: _____

Name of Representative: _____

Signature of Representative: _____

Date: _____